

Republic of the Philippines  
 City Government of Cagayan de Oro  
 BIDS AND AWARDS COMMITTEE (BAC) - GOODS  
 City Hall, Cagayan de Oro City

**REQUEST FOR QUOTATION (NEGOTIATED 53.9 SMALL-VALUE OF THE 2016 RIRR OF RA 9184)**


**Supply & Delivery 1 unit Table Top Pressure Steam Sterilizer; PR No. 21-2612 dated September 6, 2021.**

Date : October 13, 2021  
 Quotation No. : 2262-21

\_\_\_\_\_  
 Company Name

\_\_\_\_\_  
 Address

Please quote your lowest price on the item/s listed below, subject to the General Conditions on this page, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than 2:00PM on October 20, 2021 in the return envelope attached herewith.

By Authority of the BAC  
  
**TEODORO A. S. BUGA-A, JR.**  
 BAC Chairperson

- NOTE:**
- DELIVERY PERIOD WITHIN THIRTY (30) CALENDAR DAYS**
  - WARRANTY SHALL BE FOR A PERIOD OF THREE (3) MONTHS FOR SUPPLIES & MATERIALS, ONE (1) YEAR FOR EQUIPMENT, FROM DATE OF ACCEPTANCE BY THE PROCURING ENTITY;**
  - PRICE VALIDITY SHALL BE FOR A PERIOD OF 90 CALENDAR DAYS;**
  - PhilGEPS REGISTRATION CERTIFICATE/NUMBER SHALL BE ATTACHED UPON SUBMISSION OF THE QUOTATION;**
  - SUBMIT CURRENT AND VALID MAYOR'S/BUSINESS PERMIT**
  - SUPPLIERS FOR EQUIPMENT SHALL SUBMIT BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED; AND**
  - DULY NOTARIZED OMNIBUS SWORN STATEMENT PRIOR ISSUANCE OF NOTICE OF AWARDS.**

Item No.	ITEM & DESCRIPTION	BRAND & MODEL	QTY	UNIT PRICE
1	Table Top Pressure Steam Sterilizer: Brandnew		1 Unit	
	Class B Type sterilizer; Capacity: 30 liters;			
	Sterilization Temperature and Pressure :			
	Atleast 121 deg. C to 134 deg. C; Atleast Round Shape			
	Chamber and highest grade stainless steel; Safety			
	Functions such as: safety valve, overcurrent blocking			
	(NFB) Internal and external overheating sensor,			
	chamber overheating with auto door lock; With thermal			
	printer; with noise reduction (with sound level during			
	operation) Less than 75dB; with steam check valve			
	and automated power blocking system; with 6.5 liter			
	liter water tank; Dimensions: Atleast 600mm (W) x			
	670mm (L) x 430mm (H); with atleast 14 individual			
	and specific buttons parameters for easy use such			
	as: Pass Function, Re-dry function, Menu Leak Test,			
	Bowie Dick Test, Gloves, Packs Instrument, Liquids;			
	One (1) year warranty on parts and service;			
	with on standby biomedical engineer of the supplier			
	in case of emergency repairs; FREE quarterly			
	preventive maintenance; And parts replacement			
	checklist FREE Demo of the delivery unit.			

PhilGEPS Registration Number: \_\_\_\_\_

After having carefully read and accepted your General Conditions, I/we quoted you on the item at prices noted above.

\_\_\_\_\_  
 Printed Name / Signature / Designation

\_\_\_\_\_  
 Tel. No. / Cellphone No. / E-mail Address

\_\_\_\_\_  
 Date