

Republic of the Philippines
 City Government of Cagayan de Oro
 BIDS AND AWARDS COMMITTEE (BAC) - GOODS
 City Hall, Cagayan de Oro City

REQUEST FOR QUOTATION (NEGOTIATED 53.9 SMALL-VALUE OF THE 2016 RIRR OF RA 9184)


Supply and Delivery of 1 unit Automated External Defibrillator; PR No. 21-3191 dated October 13, 2021

Company Name _____

Date : December 3, 2021
 Quotation No. : 2535-21

Address _____

Please quote your lowest price on the item/s listed below, subject to the General Conditions on this page, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than 2:00PM on December 10, 2021 in the return envelope attached herewith.

By Authority of the BAC:

TEODORO A. SABUGA-A, JR.
 BAC Chairperson

- NOTE:**
1. DELIVERY PERIOD WITHIN **THIRTY (30) CALENDAR DAYS**
 2. WARRANTY SHALL BE FOR A PERIOD OF THREE (3) MONTHS FOR SUPPLIES & MATERIALS, ONE (1) YEAR FOR EQUIPMENT, FROM DATE OF ACCEPTANCE BY THE PROCURING ENTITY;
 3. PRICE VALIDITY SHALL BE FOR A PERIOD OF 90 CALENDAR DAYS;
 4. PhilGEPS REGISTRATION CERTIFICATE/NUMBER SHALL BE ATTACHED UPON SUBMISSION OF THE QUOTATION;
 5. SUBMIT CURRENT AND VALID MAYOR'S/BUSINESS PERMIT
 6. SUPPLIERS FOR EQUIPMENT SHALL SUBMIT BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED; AND
 7. DULY NOTARIZED OMNIBUS SWORN STATEMENT PRIOR ISSUANCE OF NOTICE OF AWARDS.

Item No.	ITEM & DESCRIPTION	BRAND & MODEL	QTY	UNIT PRICE
	is advised when shock is advised; Size: At least 6 cm x 18 cm x 22 cm (At least 2.4" x 7.1" x 8.9")			
	D x H x W; Weight with battery and pads case: At least 1.6 kgs. (3.5 lbs.); Patients analysis			
	Evaluates patient ECG to determine if a rhythm is shockable; Rhythms considered shockable are ventilator fibrillation(VF) and certain ventricular tachycardias (VT) associated with lack of circulation; With Sensitivity / Specificity meets with AAMI DF80 guidelines and AHA recommendations for adult defibrillation; Shock advised able to deliver a shock as soon as the device indicates a shock is advised; With Quick Shock: Able to deliver a shock after the last chestcompression of a COR interval, typically in 8 seconds; With Shock to shock cycle time: Typically less than 20 seconds between shocks in a series; With Free Ten (10) pads for the AED			
	With Status Indicators: Blinking Green light indicates ready for use; audible "chirp" indicates			

PhilGEPS Registration Number: _____

After having carefully read and accepted your General Conditions, I/we quoted you on the item at prices noted above.

Printed Name / Signature / Designation _____

Tel. No. / Cellphone No. / E-mail Address _____

Date _____

Republic of the Philippines
 City Government of Cagayan de Oro
 BIDS AND AWARDS COMMITTEE (BAC) - GOODS
 City Hall, Cagayan de Oro City

REQUEST FOR QUOTATION (NEGOTIATED 53.9 SMALL-VALUE OF THE 2016 RIRR OF RA 9184)

Supply and Delivery of 1 unit Automated External Defibrillator; PR No. 21-3191 dated October 13, 2021


Company Name _____

Date : December 3, 2021
 Quotation No. : 2535-21

Address _____

Please quote your lowest price on the item/s listed below, subject to the General Conditions on this page, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than 2:00PM on December 10, 2021 in the return envelope attached herewith.

By Authority of the BAC:


TEODORO A. SABUGA-A, JR.
 BAC Chairperson

NOTE:

1. DELIVERY PERIOD WITHIN **THIRTY (30) CALENDAR DAYS**
2. WARRANTY SHALL BE FOR A PERIOD OF THREE (3) MONTHS FOR SUPPLIES & MATERIALS, ONE (1) YEAR FOR EQUIPMENT, FROM DATE OF ACCEPTANCE BY THE PROCURING ENTITY;
3. PRICE VALIDITY SHALL BE FOR A PERIOD OF 90 CALENDAR DAYS;
4. PhilGEPS REGISTRATION CERTIFICATE/NUMBER SHALL BE ATTACHED UPON SUBMISSION OF THE QUOTATION;
5. SUBMIT CURRENT AND VALID MAYOR'S/BUSINESS PERMIT
6. SUPPLIERS FOR EQUIPMENT SHALL SUBMIT BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED; AND
7. DULY NOTARIZED OMNIBUS SWORN STATEMENT PRIOR ISSUANCE OF NOTICE OF AWARDS.

Item No.	ITEM & DESCRIPTION	BRAND & MODEL	QTY	UNIT PRICE
	need for maintenance; With Battery Insertion			
	Test: Upon battery insertion, extensive automatic self-test and user-interactive test check device			
	readiness; Pads integrity test: Specifically tests the readiness for use of pads (gel moisture);			
	With daily automatic self-test: Tests Internal circuitry, waveform delivery system, pads & battery capacity; With Artifact detection:			
	Allows ECG analysis even in the presence of most pacemaker artifact & electrical noise sources;			
	Other artifacts are detected & corrective voice prompts issued; With One (1) year Warranty			
	on Parts & Services; With Free Quarterly Maintenance; With Suppliers Standby Biomed			
	Technician in case of problems for troubleshooting; With Complete Accessories			
	and Warranty Certificates.			
	X-X-X-X-X-X-X-X-X-X-X-X-X-X			

PhilGEPS Registration Number: _____

After having carefully read and accepted your General Conditions, I/we quoted you on the item at prices noted above.

Printed Name / Signature / Designation _____

Tel. No. / Cellphone No. / E-mail Address _____

Date _____